

GOVT. MEDICAL COLLEGE & HOSPITAL, CHANDIGARH
Sector 32-B, Chandigarh-160030 Tel. 0172- 2665454-57, Fax NO. 0172-2608488

Subject: Quotation
Please quote your rate and other terms and conditions for Shredder Machine (Specifications Attached) with certificates that the rates quoted are not higher than those quoted to other departments / Institution.

Quotation for Shredder Machine due on 30.07.2010

Equipment/Machine	Quantity
Shredder Machine	1

Technical Specifications are attached for Shredder Machine
Other terms and conditions of the supply will be as under:-

1. Sales tax will be paid extra, if applicable provided it is made clear in the quotation that sale tax will be charged extra. Otherwise it will be presumed that sale tax is not to be paid extra. Similar will be in case of Excise duty.
2. The rates must be quoted for Chandigarh otherwise the quotation is likely to be rejected.
3. The rates should be quoted both in words and figures, failing which the quotation will not be considered.
4. Validity period for each item may please be specifically mentioned in your quotation (which should be atleast for one year)
5. Delivery period for each item may please be specifically mentioned in your quotation, which should not more than 30 days.
6. The quotation may be addressed to Director Principal, Govt. Medical College Hospital, Sector-32 Chandigarh.
7. A Penalty of 5% per month will be deducted for belated supplies and the event of non- supply the name of the firm will be removed from mailing list.
8. The quotation may be submitted to Deputy Medical Superintendent, 2nd Level, Block -D, GMCH-32, Chandigarh.

This may please be treated as MOST URGENT and your reply/quotation should reach this office by 30.07.2010

Deputy Medical Superintendent
Govt. Medical College & Hospital
Chandigarh

Technical Specification for Shredder Machine

Shredder Machine complete with all suitable accessories

Waste to be shredded Plastic Waste (Blood bags, round plastic pipes, syringes, I.V. Sets, plastic bottles etc) and sharps (Needles etc.)

Capacity 50-60 kg/hr

Loading From the Top

Source of Energy should operate on 440V, Three Phase Motor

Grinder

Size

- 15" complete with six Nos. blades, out of which 2 blades
- Should be rotary and 4 blades should be adjustable
- These blades should be inside and outside fitted.
- Should be able to cut hollow tubes

Blade

- Blade material EN31 with tempered
- Inside blades (2 pieces) size 15"
- Outside blade (4 pieces) size 15"
- Non Corrosive and Harden Steel
- Blades should be sharp.
- Two extra set of blades of above mentioned specification

Hopper

- Hopper should specialty fabricated from sheet 3mm
- Top feeder to avoid spillage as per model design

Housing

- The main housing should be fabricated on heavy duty mild steel plate.
- inside Rotor should complete with 2 pieces Bearing
- Bearing Size : 6313
- Filter: 1 pieces, Hole size 12mm/12.0mm
- Non rusting Material of housing & Hopper (Certificate to be issued)

Frame

Grinder should be fitted on heavy duty MS Angle Stand with tray and all moving parts should be covered with Safety guards

Motor

- Grinder should be driven by 7.5 HP AC Motor with Belt & Pulley
- Bi-directional motion to avoid jamming

Noise Pollution

Shredder should be designed and mounted, so as NOT to generate Noise and Vibration

Guarantee / Warranty Others

Guarantee / Warranty should be inclusive of comprehensive parts against any manufacturing defects for a period of two years

- Provider will do complete installation, electric fittings & foundations of machine
- To avoid rusting, should be applied rust free paint and any other material/coating

CAMC

CAMC for five years after warranty