

GOVERNMENT MEDICAL COLLEGE AND HOSPITAL, SECTOR 32, CHANDIGARH

Notice for procurement of Eagle Eye Platinum IVUS Catheters compatible with Volcano IVUS system being proprietary article under rule 166 of GFR, 2017.

The department of Physiology of this institute intends to purchase 'Eagle Eye Platinum IVUS Catheters compatible with Volcano IVUS system' having following detailed specifications:

Specification:

The catheter should have:

Frequency/type: 20 MHz digital 64 element
Vessel: Coronary, peripheral
Maximum imaging diameter: 20 mm
Diameter at transducer: 3.5F
Tip entry profile: 0.019"
Working length: 150 cm
Wire lumen length: 24 cm
Maximum guide wire: 0.014"
Proprietary imaging: VH IVUS, ChromaFlo

As per knowledge of Department of Physiology, the above said article is only manufactured by Philips, 3721 Valley Centre Drive, Suite 500 San Diego, CA 92130 USA, and the said firm sells the same through its authorized agent/ reseller M/s Horizon Healthcare Solutions, Unit no. 9, Ground Floor, U. S. Complex, 120, Mathura Road, Opp. Apollo Hospital, New Delhi, PIN-110076, in the Chandigarh region. The proprietary certificate issued by OEM is attached as annexure I.

In case, there is any other OEM for the above said article, then they are requested to submit their proposal to the Director Principal, GMCH, Chandigarh through email dpgmch-chd@gmch.gov.in / hard copy latest by 11.04.2020, failing which it will be presumed that there is no other firm who manufacture the required article/ equipment and purchase will be processed and finalized from the available source.



Dr. Anita S Malhotra,
Professor and Head
Department of Physiology,
GMCH, Sector 32,
Chandigarh



Dr. Kiran Prakash,
Assistant Professor,
Department of Physiology,
GMCH, Sector 32,
Chandigarh

PROPRIETARY ARTICLE CERTIFICATE

TO WHOM IT MAY CONCERN

This is to certify that Eagle eye Platinum IVUS Catheter which is compatible with Philips Volcano IVUS System manufactured by Phillips India Limited 8th Floor, DLF 9-B, DLF Cyber City, DLF Phase - III, Gurgaon, - 122002, India, is proprietary product of Philips Volcano. Kindly find below the specifications of Philips Volcano Eagle eye Platinum IVUS catheter.

Specifications of EEP

Frequency/Type	20 MHz Digital IUVS Catheter
Plug and play	Yes
Indication	Coronary
Maximum imaging diameter	20 mm
Shaft profile at transducer	3.5F
Distal Shaft profile	2.9F
Radiopaque markers	3 Markers 10 mm apart
Tip to transducer length	EEP-10mm / EEP ST - 2.5mm
Working Length	150cm
Wire Lumen Length	2.4cm
Maximum guide wire	0.014''
Minimum guide catheter or Sheath size	5F
Hydrophilic Coating	Yes

Yours faithfully,
M/s Philips India Ltd

Sincerely yours,
For Philips India Limited


Kamalakannan
Business manager- ISTD, Philips



Cell: 9380180019