

**GOVERNMENT MEDICAL COLLEGE & HOSPITAL, SECTOR -32, CHANDIGARH**


**Notice for procurement of DSEAK SYSTEM being Proprietary Article under Rule 166 of 2017.**


The Department of Ophthalmology of this institution intends to purchase **DSEAK System** having following detailed specification:

As per knowledge of Ophthalmology Department above said article/equipment is only manufactured by M/s Gebauer medizintechnik GmbH, Germany and the said firm sells the same through its authorized agent/reseller M/s Neo Health Care, Dehradun in the Chandigarh region. The Proprietary certificate issued by the OEM is attached as Annexure – I.

In case, there is any other, OEM for the above said article, then they are requested to submit their proposal to the Director Principal, GMCH, Chandigarh through e-mail [dpgmch-chd@gmch.gov.in](mailto:dpgmch-chd@gmch.gov.in) hard copy latest by 20.03.2019 positively, failing which it will be presumed that there is no other firm who manufacture the required article/equipment and purchase will be processed and finalized from the available source.

  
(Dr. R. K. Bansal)  
Associate Professor

  
(Dr. Subina Narang),  
Associate Professor

  
28/02/19  
(Dr. S. K. Arya)  
Prof. & Head

## Technical Specification of DSAEK System: (Proprietary Item)

1. The DSAEK System should have nasal facility on recipient during refractive procedure.
2. System should have hands free linear cutting facility during Corneal Transplantation procedure.
3. System should have pre – configured cutting set.
4. System should have hand piece with autoclave cover.
5. System should have various cutting depth ranging from 250 $\mu$ m to 600 $\mu$ m with a difference of 50 $\mu$ .
6. System should have autoclavable cable connecting to handpiece.
7. System should have Glass Applinator for measure cutting diameter before cut.
8. System should have suction ring of 19/8.5, 19/9.5, 19/10.0
9. System should have Re-lease key facility to dismantle the hand piece.
- ✓ 10. System should have EU/CE approved & certified

### Consumables:

1. cutting heads of various depths:


a) SALK Set (250 $\mu$ ) -	05 nos.
b) SALK Set (300 $\mu$ ) -	05 nos.
c) DSAEK Set (350 $\mu$ ) -	05 nos.
d) DSAEK Set (400 $\mu$ ) -	05 nos.
e) DSAEK Set (450 $\mu$ ) -	05 nos.
f) DSAEK Set (500 $\mu$ ) -	05 nos.
g) DSAEK Set (550 $\mu$ ) -	05 nos.
h) DSAEK Set (600 $\mu$ ) -	05 nos.


- ✓ 2. BD R 87 G infusions Tube 150 cm SLc with Roller lock – 05 nos.
3. Disposable vacuum Tubing SLc – 05 nos.


### Terms & Conditions:

1. Warranty for 3 years.
2. CAMC for 5 years after the expiry of warranty.

It is certified that these specifications are tailor made as per Gebauer SLc DSEAK system.

  
Dr. R. K. Bansal  
Associate Professor

  
Dr. Jitender  
Associate Professor

  
Dr. Sudesh Kumar Arya  
Prof. & Head