


GOVERNMENT MEDICAL COLLEGE & HOSPITAL, SECTOR -32, CHANDIGARH


Notice for procurement of Eye Bank Specular Microscope being Proprietary Article under Rule 166 of 2017.


The Department of Ophthalmology of this institution intends to purchase **Eye Bank Specular Microscope** having following detailed specification:

As per knowledge of Ophthalmology Department above said article/equipment is only manufactured by M/s Konan Medical Inc. 10-29, Miyanishicho, Nishinomiya-Hyogo, Japan and the said firm sells the same through its authorized agent/reseller M/S BioMedix Optotechnik & Devices Pvt. Ltd., Amarjyothi Layout – Domlur, Bangalore in the Chandigarh region. The Proprietary certificate issued by the OEM is attached as Annexure – I.

In case, there is any other, OEM for the above said article, then they are requested to submit their proposal to the Director Principal, GMCH, Chandigarh through e-mail dpgmch-chd@gmch.gov.in hard copy latest by 02.04.2019 positively, failing which it will be presumed that there is no other firm who manufacture the required article/equipment and purchase will be processed and finalized from the available source.


(Dr. R. K. Bansal)
Associate Professor


(Dr. Suresh Kumar)
Professor


14/3/19
(Dr. S. K. Arya)
Prof. & Head

Dr. R.K. BANSAL
Associate Professor
Deptt. of Ophthalmology &
Member Disability Board,
Govt. Medical College & Hospital
Sector 32-B, Chandigarh-160030

Dr. Suresh Kumar
Professor
Deptt. of Ophthalmology
Govt. Medical College & Hospital
Sector 32, Chandigarh

Dr. Sudesh K. Arya
Prof. & Head
Deptt. of Ophthalmology
Govt. Medical College &
Hospital, Chandigarh

Revised Technical Specifications of EYE Bank Specular Microscope

1	The equipment should have Ultra – wide viewing field areas of 1000 X 750µm.
2	The equipment should have a camera of CMOS Sensor with Dual capabilities for full graft imaging with 2 X zoom option for the image.
3	The equipment should have an illumination of Halogen lamp.
4	The equipment should have a reliable cellular analysis method with analytical tools such as Center method & Flex Centre method
5	The equipment should have a centre method with multi area analysis selectable upto 4 frames each of 400 X 300 µm.
6	The equipment should have a built in Pachymeter
7	The equipment should have a moving range of stage X : 16 mm, Y : 16mm, Z: 16mm with a tilt angle of 15°
8	The equipment should have a user friendly latest software with an integrated database
9	The equipment should be quoted with compatible computer & printer
10	The equipment should have original table
11	FDA/EC/BIS certification
	<p>The equipment should be supplied with the following spares (one time purchase):</p> <ul style="list-style-type: none"> a. Vial Adapter b. Vial Adapter Cap c. Vial Adapter Spare d. Vial Adapter Set e. Cushioning Box f. Chamber holder g. Chamber Cap

Terms & Conditions:


1. Warranty for 5 years.
2. CAMC for 5 years after the expiry of warranty.



(Dr. Parul Ichhpujani),
Associate Professor



(Dr. R.K. Bansal)
Associate Professor



(Dr. S.K. Arya)
Prof. & Head

Dr. Parul Ichhpujani
Associate Professor
Deptt. of Ophthalmology
Govt. Medical College &
Hospital, Chandigarh

Dr. R.K. BANSAL
Associate Professor
Deptt. of Ophthalmology &
Member Disability Board,
Govt. Medical College & Hospital
Sector 32-B, Chandigarh-160030

Dr. Sudesh K. Arya
Prof. & Head
Deptt. of Ophthalmology
Govt. Medical College &
Hospital, Chandigarh